

Agilent Purification System – Site Preparation Checklist

Thank you for purchasing an Agilent **instrument**. To get you started and to assure a successful and timely installation, please refer to this specification or set of requirements.

Correct site preparation is the key first step in ensuring that your instruments and software systems operate reliably over an extended lifetime. This document is an **information guide AND checklist** prepared for you that outlines the supplies, consumables, space and utility requirements for your equipment for your site.

Customer Responsibilities

Make sure your site meets the following prior specifications before the installation date. For details, see specific sections within this checklist, including:

- The necessary laboratory or bench space is available.
- The environmental conditions for the lab as well as laboratory gases and plumbing.
- The power requirements related to the product (e.g., number & location of electrical outlets).
- The required operating supplies necessary for the product and installation.
- Please consult Other Requirements section below for other product-specific information.

If Agilent is delivering installation and familiarization services, users of the instrument should be present throughout these services; otherwise, they will miss important operational, maintenance and safety information.

- The necessary access to the customer provided workstation is available.
- If networked, the necessary details have been reserved.

Important Customer Information

1. If you have questions or problems in providing anything described as a Customer Responsibilities above, please contact your local Agilent or partner support/service organization for assistance prior to delivery. In addition, Agilent and/or its partners reserve the right to reschedule the installation dependent upon the readiness of your laboratory.
2. Should your site not be ready for whatever reasons, please contact Agilent as soon as possible to re-arrange any services that have been purchased.
3. Other optional services such as additional training, operational qualification (OQ) and consultation for user-specific applications may also be provided at the time of installation when ordered with the system, but should be contracted separately.

Agilent Purification System – Site Preparation Checklist



Dimensions and Weight

Identify the laboratory bench space before your system arrives based on the table below.

Pay special attention to the **total height and total weight requirements for all system components you have ordered and avoid bench space with overhanging shelves**. Also pay special attention to the total weight of the modules you have ordered to ensure your laboratory bench can support this weight.

Special Notes

Product Number	Instrument Description	Weight		Height		Depth		Width	
		Kg	lbs	cm	in	cm	in	cm	in
G9300A,G9301A G9306A	Agilent 218 SDM	23.6	52.0	19.7	7.75	46.4	18.3	29.2	11.5
G9302A,G9303A	Agilent SD-1 SDM	34.0	75.0	26.7	10.5	56.0	22.0	40.6	16.0
G9309A	Agilent 325 UV/Vis Detector	15.5	34.0	21.2	8.3	47.5	18.7	29.6	11.7
G9331A,G9332A	Agilent 410 Auto Sampler	19.0	42.0	34.0	13.4	50.0	19.7	30.0	11.8
G9340A	Agilent 440 Fraction Collector	18.0	34.0	51.0	21.0	28.5	12.0	49.0	20.0



Environmental Conditions

Operating your instrument within the recommended temperature ranges insures optimum system performance and lifetime.

Special Notes

1. Performance can be affected by sources of heat & cold e.g. direct sunlight, heating/cooling from air conditioning outlets, drafts and/or vibrations.
2. The site's ambient temperature conditions must be stable for optimum performance.
3. The area selected for operation must be free from drafts, corrosive atmospheres and vibration.
4. Limit dust levels to less than 36 000 000 particles (0.5 microns or larger) per cubic meter of air. (This is equivalent to a clean office)

**Agilent Purification System -
Site Preparation Checklist**

Product Number	Instrument Description	Operating temp range °C (F)	Operating humidity range (%)
G9300A,G9301A,G9306A	Agilent 218 SDM	3-40 (41-104)	20-80
G9302A,G9303A	Agilent SD-1 SDM	4-40 (39-104)	20-80
G9309A	Agilent 325 UV/Vis Detector	10-35 (50-95)	5-95
G9331A,G9332A	Agilent 410 Auto Sampler	10-40 (50-104)	20-80
G9340A	Agilent 440 Fraction Collector	10-35 (50-95)	8-80



Power Consumption

Special Notes

1. If a computer system is supplied with your system, be sure to account for those electrical outlets.
2. Avoid using power supplies from a source that may be subject to electrical interference from other services (such as large electric motors, elevators, welders and air conditioning units).
3. Only single phase power is required.

Product Number	Instrument Description	Line Voltage & Frequency (V, Hz)	Maximum Power Consumption (VA)	Maximum Power Consumption (W)
G9300A,G9301A,G9306A	Agilent 218 SDM	100-240, 50-60	550	N/A
G9302A,G9303A	Agilent SD-1 SDM	100-240, 50-60	1000	N/A
G9309A	Agilent 325 UV/Vis Detector	100-240, 50-60	130	N/A
G9331A,G9332A	Agilent 410 Auto Sampler	115-230, 50-60	200	N/A
G9340A	Agilent 440 Fraction Collector	100-240, 50-60	100	75

**Agilent Purification System –
Site Preparation Checklist**



Required Operating Supplies by Customer

Special Notes

- For information on Agilent consumables, accessories and laboratory operating supplies, please visit <http://www.chem.agilent.com/en-US/Products/consumables/Pages/default.aspx>

Item Description (including dimensions etc)	Vendor's Part Number (if applicable)	Recommended Quantity
n/a	n/a	n/a



Other Requirements

Wiring Requirements

The Pumps, Sampler and Fraction Collector are controlled via RS232/RS422, and must be installed close to the controlling PC. The length of the flat ribbon RS 232 cable is about 1.5 m. Customer should have available HPLC grade Acetonitrile and water with a dry residue below 1 ppm or MS grade solvents.

Waste Requirements

Operation of the Agilent Purification System requires the use of a waste container for the disposal of excess fluids. Suitable tubing is supplied with the HPLC system for use with most solvents. A chemically inert container, which is appropriately sized to hold waste coming from the system must be provided by the system user. It should be located underneath the workbench where it is protected by the bench and in full view of the operator.

Network Requirements

The Agilent 325 UV/vis Dual WL detector requires an Ethernet connection to the PC via a shielded twisted pair Ethernet crossover cable Category 5 or better (provided with the detector). If connection is required to an Ethernet network, then a shielded twisted pair Ethernet non-crossover cable Category 5 or better will be required (not supplied). Most PC's come pre-configured with an Ethernet connection either built into the motherboard or with an Ethernet network card installed. Check that the intended PC has this configuration and if not, a network card will have to be fitted. Refer to the installation chapter for details.

Agilent Purification System - Site Preparation Checklist

If the user is supplying their own PC, then they are responsible for installing and configuring the card. They are also responsible for setting up and maintaining any LAN configuration where a detector may be used.

All network issues are to be dealt with by the users.

Important Customer Web Links

- For additional information about our solutions, please visit our web site at <http://www.chem.agilent.com/en-US/Pages/HomePage.aspx>
- Need to get information on your product?
Literature Library - <http://www.agilent.com/chem/library>
- Need to know more?
Customer Education - <http://www.agilent.com/chem/education>
- Need technical support, FAQs? - <http://www.agilent.com/chem/techsupp>
- Need supplies? - <http://www.agilent.com/chem/supplies>